


sira
Certification Service



ENVIRONMENT
AGENCY

PRODUCT CONFORMITY CERTIFICATE

This is certify that the

4500 Premier Dust Monitor

manufactured by:

Land Instruments International Ltd

Stubley Lane
Dronfield
Derbyshire
S18 1DJ
England

has been assessed by Sira Certification Service
and found to comply with:

**MCERTS Performance Standards for Continuous Emission
Monitoring Systems, Version 2, Revision 1 (April 2003)**

Certification Ranges :

Optical Density (Extinction)	0 - 0.1 to 0 - 3.0
Opacity	0 - 10 % to 0 - 100 %
Dust Concentration	0 - 15 mg/m ³ (Min pathlength 5 metres) to 0 - 3000 mg/m ³

* 0 - 0.1 extinction translates to 0 - 15mg/m³ for a minimum pathlength of 5 metres.

Certification is awarded in respect of the conditions stated in this certificate

Project No:	674/0162
Certificate No:	Sira MC 040029/00
Initial Certification:	01 st July 2004
This Certificate Issued:	01 st July 2004
Renewal Date:	30 th June 2009


Chief Executive

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

South Hill, Chislehurst, Kent, BR7 5EH, England

Tel: 020-8467-2636 Fax: 020-8468 1841

This certificate may only be reproduced in its entirety and without change