



National Accreditation Board for Testing and Calibration Laboratories

CERTIFICATE OF ACCREDITATION

AMETEK INSTRUMENTS INDIA PRIVATE LTD.

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2017

**"General Requirements for the Competence of Testing &
Calibration Laboratories"**

for its facilities at

**#148, DIVYASREE NR ENCLAVE, BLOCK-A & B,, 4TH FLOOR, SITE#1, EPIP INDUSTRIAL AREA,
WHITEFIELD, BENGALURU, KARNATAKA, INDIA**

in the field of

CALIBRATION

Certificate Number: **CC-2041**

Issue Date: **31/03/2021**

Valid Until:

30/03/2023

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued
satisfactory compliance to the above standard & the relevant requirements of NABL.
(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Name of Legal Identity : **Ametek Instruments India Private Limited**

Signed for and on behalf of NABL



N. Venkateswaran
Chief Executive Officer



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

AMETEK INSTRUMENTS INDIA PRIVATE LTD., #148, DIVYASREE NR ENCLAVE, BLOCK-A & B,, 4TH FLOOR, SITE#1, EPIP INDUSTRIAL AREA, WHITEFIELD, BENGALURU, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2041

Page No

1 of 3

Validity

31/03/2021 to 30/03/2023

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using RTD sensor with Temperature Indicator,IR Thermometer, Black Body Furnaces of emissivity =0.995 by Comparison Method	0 °C to 300 °C	1.24°C
2	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using IR Thermometer, Black Body Furnaces by Comparison Method, Emissivity is 0.998	1000 °C to 1200 °C	2.96°C
3	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using IR Thermometer, Black Body Furnaces by Comparison Method, Emissivity is 0.998	1200 °C to 1400 °C	3.45°C
4	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using IR Thermometer, Black Body Furnaces by Comparison Method, Emissivity is 0.998	1400 °C to 1500 °C	3.5°C
5	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using IR Thermometer, Black Body Furnaces by Comparison Method, Emissivity is 0.998	1500 °C to 1600 °C	3.71°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

AMETEK INSTRUMENTS INDIA PRIVATE LTD., #148, DIVYASREE NR ENCLAVE, BLOCK-A & B,, 4TH FLOOR, SITE#1, EPIP INDUSTRIAL AREA, WHITEFIELD, BENGALURU, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2041

Page No

2 of 3

Validity

31/03/2021 to 30/03/2023

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
6	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using RTD sensor with Temperature Indicator,IR Thermometer, Black Body Furnaces of emissivity =0.995 by Comparison Method	300 °C to 550 °C	1.25°C
7	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using IR Thermometer, Black Body Furnaces by Comparison Method, Emissivity is 0.998	550 °C to 1000 °C	2.41°C



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

AMETEK INSTRUMENTS INDIA PRIVATE LTD., #148, DIVYASREE NR ENCLAVE, BLOCK-A & B,, 4TH FLOOR, SITE#1, EPIP INDUSTRIAL AREA, WHITEFIELD, BENGALURU, KARNATAKA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-2041

Page No

3 of 3

Validity

31/03/2021 to 30/03/2023

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Site Facility					
1	THERMAL-TEMPERATURE	Non Contact IR Pyrometer /thermometer	Using IR Thermometer, Black Body Furnaces of Emissivity=0.98.00 by Comparison Method	0°C to 150°C	1.0 °C
2	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using IR Thermometer, Black Body Furnaces of Emissivity= 0.98 by Comparison Method	1000 °C to 1200 °C	3.39°C
3	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using IR Thermometer, Black Body Furnaces of Emissivity =0.98 by Comparison Method	150 °C to 400 °C	2.7°C
4	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using IR Thermometer, Black Body Furnaces at Emissivity = 0.98 by Comparison Method	400 °C to 700 °C	2.54°C
5	THERMAL-TEMPERATURE	Non Contact IR Pyrometer/thermometer	Using IR Thermometer, Black Body Furnaces by Comparison Method, Emissivity is 0.98	700 °C to 1000 °C	2.8°C

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.